The sensitivity of the system falls from 91 to 86% when 6-h validation was performed. BIRRING *et al.* [1] state that the reduction in sensitivity is most likely due to a more robust measure of true cough frequency used for the 2-h recordings. An alternative explanation is that the 2-h validation is poorly representative of the performance of the device over longer time periods. The data is presented for 24-h recordings on 50 subjects. So where are the data on sensitivity and specificity validated over a full 24 h? Previous publications have manually counted cough over 24-h periods in larger numbers of patients [2].

Furthermore, the title of the paper describes the Leicester Cough Monitor as an automated cough detection system. In fact, operator input is required to classify sounds that might have been wrongly identified as cough events in the first phase. Therefore, the system would be more accurately described as semi-automated.

This is now the third publication on the Leicester Cough Monitor without validation beyond 6 h, yet the authors invite others to contact them to use the monitor based on this data. We ask S.S. Birring and co-workers to state candidly whether this device has or has not been validated over 24 h. If it has, can they please provide the missing data on 24-h validation so that we can assess the true accuracy of the Leicester Cough Monitor?

K. McGuinness*, A. Morice#, A. Woodcock and J. Smith*

*Respiratory Research Group, University of Manchester, Manchester, #Cardiovascular and Respiratory Studies, University of Hull, Hull, UK.

STATEMENT OF INTEREST

Statements for interest for all authors can be found at www.erj.ersjournals.com/misc/statements.shtml

REFERENCES

- 1 Birring SS, Fleming T, Matos S, Raj AA, Evans DH, Pavord ID. The Leicester Cough Monitor: preliminary validation of an automated cough detection system in chronic cough. *Eur Respir J* 2008; 31: 1013–1018.
- 2 Decalmer S, Webster D, Kelsall A, McGuinness K, Woodcock A, Smith JA. Chronic cough: how do cough reflex sensitivity and subjective assessments correlate with objective cough counts during ambulatory monitoring? *Thorax* 2007; 62: 329–334.

DOI: 10.1183/09031936.00052008

From the authors:

K. MCGUINNESS and colleagues request further details regarding the validation of the Leicester Cough Monitor (LCM) over 24 h. This was not possible during our study because of limitations in battery life [1]. Therefore, we compared automated and manual cough counts during the day, at a time when the frequency of cough and other sounds was highest and the demands on the system were most stringent.

Since our paper has been published [1], we are very grateful to Professor A. Monce for providing us with the opportunity to independently validate 24-h recordings obtained from his

institution. Our findings are consistent with earlier reports suggesting that the LCM is an accurate measure of cough frequency [1–4]. Seven patients with stable unexplained chronic cough (six females, mean age 51 yrs) were randomly recruited from Castle Hill Hospital (Hull, UK). Cough frequency measured by automated LCM analysis was compared with that measured manually by one observer. All patients underwent 24-h continuous ambulatory sound recording in their own environment with a Sony Minidisc (MZ-RH1) recorder (Sony Corporation, Tokyo, Japan) and an external cardioid stereo microphone (SP-CMC-2A; The Sound Professionals, Hainesport, NJ, USA) [5]. An experienced independent observer from Castle Hill Hospital, blinded to the automated cough analysis, measured cough counts manually by audio and visual inspection. Unfiltered sound recordings were also analysed using the automated LCM by an observer blinded to the manual cough counts [1]. Coughs were identified as single events whether they occurred as isolated events or in a cluster. All patients gave written consent and the local ethics committee approved the study. There was no significant difference between manual and LCM automated cough counts (mean $\pm$ SEM 23 ± 7 versus 24 ± 6 coughs $\cdot$ patient $^{-1} \cdot$ h $^{-1}$; mean difference = 1; $p = 0.5$) and there was a highly significant correlation between manual and automated coughs $\cdot$ patient $^{-1} \cdot$ h $^{-1}$ ($r = 0.97$, $p < 0.001$). The intra-class correlation coefficient of agreement between manual and automated cough counts was also highly significant (0.98, $p < 0.001$; fig. 1). We were unable to determine sensitivity analysis or to carry out an hour-by-hour analysis because of the difficulty in synchronising time stamping of detected events between the two cough recording systems. However, the close agreement between the numbers of coughs detected in the 24-h recordings across a wide range of cough frequencies increases confidence that the LCM has high sensitivity and specificity for detecting coughs and provides further validation of our system against independently acquired and counted recording.

Automated cough detection systems are potentially an important advance in an important and grossly under

FIGURE 1. Bland-Altman plot comparing cough frequency measured by an automated cough monitor (Leicester Cough Monitor) and manually by an observer.: mean difference between manual and automated cough counts, —: 95% limits of agreement.

researched area [6]. To date, our validation of the Leicester Cough Monitor is the most extensive for cough monitoring systems. We are aware of fully automated cough detection systems where the same level of validation is not available [7]. We encourage our colleagues to publish similar validation data so that independent researchers are able to compare the merits of the available system.

S.S. Birring*, V.M. Mann[#], S. Matos[¶], N. Yousaf⁺, C. Wood*, T. Fleming*, D.H. Evans[¶], A.H. Morice[#] and I.D. Pavord⁺

*Dept of Respiratory Medicine, King's College Hospital, London, [#]Cardiovascular and Respiratory Studies, Hull York Medical School, Hull, [¶]Dept of Medical Physics, Leicester Royal Infirmary, and ⁺Institute for Lung Health, Dept of Respiratory Medicine, Glenfield Hospital, Leicester, UK.

STATEMENT OF INTEREST

None declared.

REFERENCES

- 1 Birring SS, Fleming T, Matos S, Raj AA, Evans DH, Pavord ID. The Leicester Cough Monitor: preliminary validation of an automated cough detection system in patients with chronic cough. *Eur Respir J* 2008; 31: 1013–1018.
- 2 Birring SS, Matos S, Patel RB, Prudon B, Evans DH, Pavord ID. Cough frequency, cough sensitivity and health status in patients with chronic cough. *Respir Med* 2006; 100: 1105–1109.
- 3 Matos S, Birring SS, Pavord ID, Evans DH. An automated system for 24-h monitoring of cough frequency: the Leicester cough monitor. *IEEE Trans Biomed Eng* 2007; 54: 1472–1479.
- 4 Matos S, Birring SS, Pavord ID, Evans DH. Detection of cough signals in continuous audio recordings using hidden Markov models. *IEEE Trans Biomed Eng* 2006; 53: 1078–1083.
- 5 Barry SJ, Dane AD, Morice AH, Walmsley AD. The automatic recognition and counting of cough. *Cough* 2006; 2: 8.
- 6 Pavord ID, Chung KF. Management of chronic cough. *Lancet* 2008; 371: 1375–1384.
- 7 Coyle MA, Keenan DB, Henderson LS, *et al.* Evaluation of an ambulatory system for the quantification of cough frequency in patients with chronic obstructive pulmonary disease. *Cough* 2005; 1: 3.

DOI: 10.1183/09031936.00060808

Local immunodiagnosis of pulmonary TB: ELISPOT or flow cytometry, PPD or ESAT-6?

To the Editors:

Immunodiagnosis is a significant addition to the diagnostic armamentarium in tuberculosis (TB) infection and disease. For some time it has been known that cells secreting type-1 cytokines, such as interferon (IFN)- γ and tumour necrosis factor- α , selectively accumulate in the lung in higher frequencies than in the blood [1], as would be expected in a disease dominated by pulmonary pathology.

In a recent issue of the *European Respiratory Journal*, JAFARI *et al.* [2] detailed the TB-specific response in bronchoalveolar lavage (BAL) and blood from 12 patients with culture-positive but smear negative disease. JAFARI *et al.* [2] used a combination of flow cytometry to determine the phenotype of the CD4 and CD8 lymphocytes concentrating on memory and activation markers, and enzyme-linked immunospot (ELISPOT), to measure the response to the TB-specific proteins early secretory antigenic target (ESAT)-6 and culture filtrate protein (CFP)-10. This study is similar in design and findings to another paper by JAFARI *et al.* [3].

One of the disadvantages of ELISPOT when compared with flow cytometry is that it cannot identify the phenotype of the cells secreting the measured cytokines (usually IFN- γ). Hence in the study by JAFARI *et al.* [2] it is through association, rather than direct measurement, that the TB-specific lymphocytes are

characterised as predominantly memory phenotype, as evidenced by CD45RO expression. Indeed the percentage of CD45RO⁺ CD4⁺ lymphocytes in TB (50%) and the control group (35%) appeared to be surprisingly low. True memory CD4⁺ lymphocytes, determined by the absence of both CD45RA and CD27 expression, were 93% in BAL in 46 controls and patients with a variety of different pathologies, compared with 23% in blood samples [4].

In the study by JAFARI *et al.* [2], the region of difference-1 antigens, ESAT-6 and CFP-10, were used as the stimulatory antigens. JAFARI *et al.* [2] noted that the TB-specific responses to these peptides in BAL were ~10 times those in peripheral blood. However, ESAT-6 and CFP-10 only encompass a small part of the TB genome and larger proteins, such as purified protein derivative (PPD), have been shown to generate much larger responses. For example, in a previous study [5] the median CD4 T cell IFN- γ response to PPD in BAL was 24% but only 0.07% in blood, a 300-fold difference. More importantly, ESAT-6 responses in BAL in a large cohort of patients were negative in 19% of cases with culture confirmed TB [6].

Most enthusiasts for using region of difference-1 antigens in TB immunodiagnosis point out the difficulties of using PPD on blood samples when patients have received the bacilli Calmette–Guerin (BCG) vaccination. However, BCG vaccination has no influence on lung antigen-specific responses when PPD is used